


## Transport Monitor



### Transport Monitor

Weight	0.95 kg (2.1 lbs) (Standard parameters with battery) 1.17 kg (2.6 lbs) (Standard parameters with internal CO2 module and battery)
Size	150x103x81 mm (5.9" x 4" x 3.2")
Type	Medical-grade color TFT LCD, capacitive touch screen, with Corning® Gorilla® Glass, support multi-touch operation.
Size & Resolution	5.5-inch, 1280 x 720 pixels (WXGA)
Waveforms	5 traces, up to 13 waveforms
External display	Medical-grade color TFT LCD, capacitive touch screen, 19-inch, 1280 x 720 pixels Up to 8 traces

### ECG

Meet standards of IEC 60601-2-27 and IEC 60601-2-25.

Lead Sets	Automatic 3/5/6/12 - lead recognition
3-lead:	I, II, III
5-lead:	I, II, III, aVR, aVL, aVF, V
6-lead:	I, II, III, aVR, aVL, aVF, Va, Vb
12-lead:	I, II, III, aVR, aVL, aVF, V1 to V6
Sweep Speed	6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
Gain Selection	x 0.125, x 0.25, x 0.5, x 1, x 2, x 4, auto
Waveform format	Standard, Cabrera
Input Signal Range	± 8 mV (p-p)
Electrode Offset Potential Tolerance	± 500 mV
Bandwidth	
Diagnostic Mode:	0.05 to 150 Hz
Monitor Mode:	0.5 to 40 Hz
Surgical Mode:	1 to 20 Hz
ST Mode:	0.05 to 40 Hz
High Freq Cut-off (for 12-lead ECG analysis):	350 Hz, 150 Hz, 35 Hz, 20 Hz selectable

### CMRR

Diagnostic:	> 90 dB
Monitor, Surgical, ST mode:	> 105 dB (with notch filter on)
Pace detection	
Amplitude:	± 2 mV to ± 700 mV
Width:	0.1 to 2 ms
Rise time:	10 to 100 µs (without overshoot)
Defibrillator Protection	Withstand 5000VAC (360J) defibrillation
Defib. Recovery Time	≤ 5 seconds
ESU recovery time	≤ 10 s
Provides Glasgow resting 12-lead ECG algorithm.	
Provides Mindray Multi(4)-lead ECG monitoring analysis algorithm.	

### Measurement Range

Adult:	15 to 300 bpm
Pediatric/Neonate:	15 to 350 bpm

Accuracy	± 1 bpm or ± 1%, whichever is greater.
Resolution	1 bpm
Arrhythmias	Adult/Pediatric/Neonate.
Monitored Arrhythmias	Asystole, VFib/VTac, VTac, Vent. Brady, Extreme Tachy, Extreme Brady, Vrrhythm, PVCs/min, Pauses/min, Couplet, Bigeminy, Trigeminy, R on T, Run PVCs, PVC, Tachy, Brady, Missed Beats, PNP, PNC, Multif. PVC, Nonsus. VTac, Pause, Irr. Rhythm, AFib.

### ECG m An T

Patient	Adult/Pediatric.
Range	- 2.0 to + 2.0 mV (RTI)
Accuracy	± 0.02 mV or ± 10%, whichever is greater (- 0.8 to + 0.8 mV)
Resolution	0.01 mV

### ECG An T

Patient	Adult/Pediatric/Neonate.
Parameters	QT, QTc, ΔQTc
QTc Formula	Bazett, Fridericia, Framingham, or Hodges
Range	
QT/QTc:	200 to 800 ms
QT-HR:	Adult: 15 to 150 bpm Pediatric/Neonate: 15 to 180 bpm
QT Accuracy	± 30 ms
Resolution	QT 4 ms; QTc 1 ms

### ECG p.r . on

Range	0 to 200 bpm
Resolution	1 rpm
Apnea Alarm Time	10, 15, 20, 25, 30, 35, 40 sec
Accuracy	
0 - 120 rpm:	± 1 rpm
121 - 200 rpm:	± 2 rpm
Lead	I, II, or auto (default: lead II)

### ECG T . vr

Meet standards of ISO 80601-2-61.

Module	Mindray, Masimo, Nellcor
Range	0 to 100 %
Resolution	1%
Accuracy	
Mindray/Nellcor:	± 2 % (70 to 100%, Adult/Pediatric); ± 3 % (70 to 100%, Neonate) Unspecified (0 to 69%)
Masimo:	± 2 % (70 to 100%, Adult/Pediatric, non-motion) ± 3 % (70 to 100%, Neonate, non-motion) ± 3 % (70 to 100%, motion) Unspecified (0 to 69%)
Perfusion indicator (PI)	Yes, for Mindray/Masimo SpO2
Pitch Tone	Yes
Dual-SpO2	Yes, SpO2, SpO2b, ΔSpO2

Pulse Rate Range		(excluding sensor error)
Mindray/Nellcor:	20 to 300 bpm	Sensitivity
Masimo:	25 to 240 bpm	5 $\mu$ V/V/mmHg
Pulse Rate Accuracy		Impedance Range
Mindray:	$\pm$ 3 bpm (20 - 300 bpm)	300 to 3000 $\Omega$
Nellcor:	$\pm$ 3 bpm (20 - 250 bpm)	PPV Range
Masimo:	$\pm$ 3 bpm (non-motion)	0 to 50 %
	$\pm$ 5 bpm (motion)	PAWP
PR Refresh Rate	1 sec	Yes
		ICP measurement
		Support
		Support waveforms overlapping.
		Pulse Rate Range
		25 to 350 bpm
		Pulse Rate Accuracy
		$\pm$ 1 bpm or $\pm$ 1 %, whichever is greater

### Pressure

Meet standard of ISO 80601-2-56.

Method	Thermal resistance
Channels	Up to 8 channels
Units of Measure	Selectable $^{\circ}$ C or $^{\circ}$ F
Range	0 to 50 $^{\circ}$ C / 32 to 122 $^{\circ}$ F
Resolution	0.1 $^{\circ}$ C, 0.1 $^{\circ}$ F
Accuracy	$\pm$ 0.1 $^{\circ}$ C or $\pm$ 0.2 $^{\circ}$ F (without probe)
Refresh Rate	1 sec

### Non-Invasive Blood Pressure

Meet standards of ISO 80601-2-30.

Method	Oscillometry
Modes	Manual, Auto, STAT, Sequence
Units of Measure	mmHg, kPa (user-selectable)
Resolution	1 mmHg

#### Systolic range

Adult:	25 to 290 mmHg
Pediatric:	25 to 240 mmHg
Neonate:	25 to 140 mmHg

#### Diastolic range

Adult:	10 to 250 mmHg
Pediatric:	10 to 200 mmHg
Neonate:	10 to 115 mmHg

#### Mean range

Adult:	15 to 260 mmHg
Pediatric:	15 to 215 mmHg
Neonate:	15 to 125 mmHg

#### Accuracy

Max Mean Error:	$\pm$ 5 mmHg
Max Standard Deviation:	8 mmHg

#### Cuff Deflation Technique

Step bleed

#### Initial Cuff Inflation

Adult:	80 to 280 mmHg (default: 160 mmHg)
Pediatric:	80 to 210 mmHg (default: 140 mmHg)
Neonate:	60 to 140 mmHg (default: 90 mmHg)

#### Over Pressure Protection

Adult/ Pediatric:	297 $\pm$ 3 mmHg
Neonate:	147 $\pm$ 3 mmHg

#### Max Measurement time

Adult/Pediatric:	180 sec
Neonate:	90 sec

#### Assisting Venous Puncture

Yes

Pulse Rate Range	30 to 300 bpm
Pulse Rate Accuracy	$\pm$ 3 bpm or $\pm$ 3 %, whichever is greater

### B

Meet standard of IEC 60601-2-34.

Number	Up to 8 channels
Measurement Range	-50 to 360 mmHg
Resolution	1 mmHg
Accuracy	$\pm$ 1 mmHg or $\pm$ 2 %, whichever is greater

### CC

Parameters	Measurement Range	Coefficient of Variation
CCO	0.25 to 25.0 L/min	$\leq$ 2%
C.O.	0.25 to 25.0 L/min	$\leq$ 2%
GEDV	40 to 4800 ml	$\leq$ 3%
SV	1 to 250 ml	$\leq$ 2%
EVLW	10 to 5000 ml	$\leq$ 6%
ITBV	50 to 6000 ml	$\leq$ 3%

(Coefficient of variation is measured using synthetic and/or database wave forms (laboratory testing.) Coefficient of variation= SD/mean error.)

TB Range	23 to 43 $^{\circ}$ C / 73.4 to 109.4 $^{\circ}$ F
TB, TI Accuracy	$\pm$ 0.1 $^{\circ}$ C (without sensor)
TB, TI Resolution	0.1 $^{\circ}$ C
pArt/pCVP Range	-50 to 300 mmHg
pArt/pCVP Accuracy	$\pm$ 1 mmHg or $\pm$ 2 %, whichever is greater

### Non-T. or C.

Meet standard of ISO 80601-2-55.

Patient	Adult/Pediatric/Neonate.
Measurement Range	0 to 150 mmHg
CO <sub>2</sub> Accuracy	
0 to 40 mmHg:	$\pm$ 2mmHg
41 to 76 mmHg:	$\pm$ 5% of reading
77 to 99 mmHg:	$\pm$ 10% of reading
100 to 150 mmHg:	$\pm$ (3 mmHg+8% of reading)
Sample Flow Rate	50 ml/min
Sample Flow Rate Tolerance	$\pm$ 15 ml/min or $\pm$ 15 %, whichever is greater.
Sweep speed	3 mm/sec, 6.25 mm/sec, 12.5 mm/sec, 25 mm/sec, 50 mm/sec
awRR range	0 to 150 rpm
awRR accuracy	
0 to 60 rpm:	$\pm$ 1 rpm
61 to 150 rpm:	$\pm$ 2 rpm
Apnea time	10, 15, 20, 25, 30, 35, 40 sec

### Ar or C.

Meet standard of ISO 80601-2-55.

Measurement Range	
etCO <sub>2</sub> :	0 to 150 mmHg
O <sub>2</sub> (optional) :	0 to 100 %
CO <sub>2</sub> Accuracy	
0 to 40 mmHg:	$\pm$ 2mmHg
41 to 76 mmHg:	$\pm$ 5% of reading
77 to 99 mmHg:	$\pm$ 10% of reading
100 to 150 mmHg:	$\pm$ (3 mmHg+8% of reading)
O <sub>2</sub> Accuracy	
0 to 25 %:	$\pm$ 1 %
25.1 to 80 %:	$\pm$ 2 %
80.1 to 100 %:	$\pm$ 3 %

Resolution	
etCO <sub>2</sub> :	1 mmHg
O <sub>2</sub> (optional) :	1 %
Sample Flow Rate	
Adult/Pediatric:	120 ml/min (with or without O <sub>2</sub> monitoring)
Neonate:	70 ml/min or 90 ml/min, selectable 90 ml/min (with O <sub>2</sub> monitoring)
Sample Flow Rate Tolerance	±15 ml/min or ±15 %, whichever is greater.
Warm-up Time	90 sec (maximum), 20 sec (typically)
Measured with a neonatal watertrap and 2.5-meter neonatal sampling line, or an adult watertrap and a 2.5-meter adult sampling line:	
Rise Time	
etCO <sub>2</sub> :	≤ 250 ms @ 70 ml/min (Neonate watertrap) ≤ 250 ms @ 90 ml/min (Neonate watertrap) ≤ 300 ms @ 120 ml/min (Adult watertrap)
O <sub>2</sub> (optional) :	≤ 800 ms @ 90 ml/min (Neonate watertrap) ≤ 750 ms @ 120 ml/min (Adult watertrap)
Sampling Delay Time	
etCO <sub>2</sub> :	≤ 5.0 sec @ 70 ml/min (Neonate watertrap) ≤ 4.5 sec @ 90 ml/min (Neonate watertrap) ≤ 5.0 sec @ 120 ml/min (Adult watertrap)
O <sub>2</sub> (optional) :	≤ 4.5 sec @ 90 ml/min (Neonate watertrap) ≤ 5.0 sec @ 120 ml/min (Adult watertrap)
awRR Range	0 to 150 rpm
awRR Accuracy	
0 to 60 rpm:	± 1 rpm
61 to 150 rpm:	± 2 rpm
Apnea Time	10, 15, 20, 25, 30, 35, 40 sec

### Respiratory CO<sub>2</sub>

Measurement Range	0 to 99 mmHg
Resolution	1 mmHg
Accuracy	
0 to 38 mmHg:	±2 mmHg
39 to 99 mmHg:	± 5 % + 0.08 % of the reading – 38 mmHg
Sample Flow Rate	50 <sup>-7.5</sup> <sub>+15</sub> ml/min
Start-up Time	30 sec (typical)
Response Time	2.9 s (typical)
awRR Range	0 to 150 rpm
awRR Accuracy	
0 to 70 rpm:	±1 rpm
71 to 120 rpm:	±2 rpm
121 to 150 rpm:	±3 rpm
Apnea time	10, 15, 20, 25, 30, 35, 40 sec

### Clinical CO<sub>2</sub>

Measurement Range	0 to 150 mmHg
Resolution	1 mmHg
Accuracy	
0 to 40 mmHg:	±2mmHg
41 to 70 mmHg:	± 5% of reading
71 to 100 mmHg:	± 8% of reading
101 to 150 mmHg:	± 10% of reading
Rise time	< 60 msec
awRR Range	0 to 150 rpm
awRR Accuracy	±1 rpm

### Diagnosis

Trends Data	> 120 hrs @ 1 min, 4 hrs @ 5 sec.
-------------	-----------------------------------

Events	1000 events, including parameter alarms, arrhythmia events, technical alarms, and so on.
NIBP	1000 sets
Interpretation of resting 12-lead ECG results	20 sets
Full disclosure	48 hours for all parameters and waveforms (8G storage card) 48 hours at maximum. The specific storage time depends on the waveforms stored and the number of stored waveforms. (2G storage card)
OxyCRG <sup>1</sup>	48 hrs
Minitrend <sup>1</sup>	Yes

### Alarm

Audible indicator	Yes, 3 different alarm tones, and prompt tone
Visible indicator	Red/yellow/cyan LED, and alarm message

### Application

Clinical Assistive Application (CAA):	ST Graphic™, BoA Dashboard™, EWS, GCS
Support calculations (drug, hemodynamic, Oxygenation, Ventilation, Renal), and Titration table.	
Support nView remote display tool	

### Wireless LAN

Protocol	IEEE 802.11a/b/g/n
Modulation Mode	DSSS and OFDM
Operating Frequency	
IEEE 802.11b/g/n (2.4G):	
ETSI/FCC/KC:	2.4 to 2.483 GHz
MIC:	2.4 to 2.495 GHz
IEEE 802.11a/n (5G):	
ETSI:	5.15 to 5.35 GHz, 5.47 to 5.725 GHz
FCC:	5.15 to 5.35 GHz, 5.725 to 5.82 GHz
MIC:	5.15 to 5.35 GHz
KC:	5.15 to 5.35 GHz, 5.47 to 5.725 GHz, 5.725 to 5.82 GHz

Channel Spacing	5 MHz @ 2.4 GHz (802.11 b/g/n) 20 MHz @ 5 GHz (802.11 a/n)
-----------------	---

Wireless Baud Rate	IEEE 802.11a: 6 to 54 Mbps IEEE 802.11b: 1 to 11 Mbps IEEE 802.11g: 6 to 54 Mbps IEEE 802.11n: 6.5 to 72.2 Mbps
--------------------	--

Output Power	< 20dBm (CE requirement: detection mode- RMS) < 30dBm (FCC requirement, detection mode- peak power)
--------------	--

Operating Mode	Infrastructure
Data Security	WPA-PSK, WPA2-PSK, WPA-Enterprise, WPA2-Enterprise (EAP-FAST, EAP-TLS, EAP-TTLS, PEAP-GTC, PEAP-MSCHAPv2, PEAP-TLS, LEAP) Encryption: TKIP and AES

### Protection

Auxiliary Output Standard	Meets the requirements of ANSI/AAMI/IEC 60601-1 for short-circuit protection and leakage current
---------------------------	--

### ECG Analog Output

Bandwidth (- 3 dB; reference frequency: 10 Hz)	
Diagnostic Mode:	0.05 to 150 Hz
Monitor Mode:	0.5 to 40 Hz

